

Quality Assurance Testing of Gamma Camera and SPECT Systems

Sharon L. White

University of Alabama at Birmingham

July 28, 2009


Disclaimers

Gamma camera images and photographs of equipment are not intended to advertise or endorse any particular manufacturer or vendor.

Outline

- Basics of Gamma Camera Operation
- Gamma Camera Calibrations
- Routine QC Tests
- SPECT Phantom Imaging
- Annual Physics Tests

Gamma Camera Operation



Position Determination


- The point where the gamma ray hits the crystal is determined by a weighted average of the signals from the group of PMTs receiving light from that event.
- The collimator localizes the origin of the gamma ray as somewhere along a specific line through the patient, since only gamma rays traveling parallel to the holes will go through. (Except for occasional septal penetration.)

Spatial Resolution

- Intrinsic resolution (R_{int}) refers to how well the crystal and PMT system localize an interaction in the crystal. Affected by crystal thickness, gamma ray energy, scatter in crystal.
- Collimator resolution (R_{coll}) refers to how well the collimator localizes the gamma ray source in the patient, affected by hole diameter and length, distance from collimator to patient.
- System resolution (R_{sys}) is combination of intrinsic and collimator resolution:

$$R_{sys} = \sqrt{R_{int}^2 + R_{coll}^2}$$

Collimator Resolution



d = hole diameter

L = hole length

X = distance from collimator
to source

Collimator Resolution

$$R_{coll} \approx \frac{d}{L} (L + x)$$


Collimators

- Parallel hole collimators used the most
- Different collimators available for different energy radionuclides – medium energy for ^{111}In and ^{67}Ga , high energy for ^{131}I
- Also have different choices for favoring high resolution vs. high sensitivity

Collimator Specifications


Type	Hole Diameter (mm)	Septal Thickness (mm)	Hole Length (mm)	Coll. Res. At 10 cm (mm)	System Res at 10 cm (mm) 9.5 mm crystal
LEGP	1.40	0.180	24.7	8.0	8.8
LEHR	1.40	0.152	32.8	6.3	7.4
MEGP	2.95	1.143	48.0	10.7	11.3
HEGP	3.81	1.727	60.0	12.0	12.5
HEHR	3.06	1.95	60.0	9.6	10.4

SPECT Operation




Camera heads rotate around patient, acquiring a set of projection images that are reconstructed into slices


SPECT Brain Projection Images




SPECT Brain Reconstructed Slices



Sagittal Slices



Coronal Slices




Gamma Camera Calibrations


- PMT gains must be balanced
- Correction Tables:
 - Energy
 - Linearity
 - Uniformity (Flood)
- Center of Rotation (COR) offset calibration for SPECT-capable cameras.

Correction Tables


No
corrections




Energy
And
Linearity



Energy
only





Energy,
Linearity,
Uniformity
(all corrections)



^{99m}Tc Intrinsic Flood Images



Correction Tables

Energy
only



Energy and
Uniformity,
No Linearity
Correction


Energy
And
Linearity
Only



All Corrections
On

^{99m}Tc Intrinsic

Intrinsic Bars – Linearity Correction Off



Routine QC Tests

- Uniformity - daily
- Spatial Resolution and Linearity - weekly
- Photopeak energy - daily
- Center of Rotation offset if camera used for SPECT – monthly or as recommended by manufacturer

Routine QC - Uniformity

- Uniformity must be checked every day that gamma camera is used, before the first patient
- Uniformity flood image may be acquired with collimator on for system (extrinsic) uniformity or collimator off for intrinsic uniformity
- 5 million counts is adequate for daily QC for large FOV camera
- Check that photopeak is centered in energy window

System Uniformity

With collimator on, use planar sheet source:



^{57}Co sheet source
10-15 mCi when new
122 keV γ
Half life 270 days



Water filled sheet source
Add 10-15 mCi $^{99\text{m}}\text{Tc}$
140 keV γ
Half life 6 hours

Intrinsic Uniformity

- General method – use $\sim 500 \mu\text{Ci } ^{99\text{m}}\text{Tc}$ point source, placed at a distance of five times the length of the camera field of view
- Some cameras have a special source holder and vendor specific procedure

Flood Images

QC 2 16 07 FORTE 1

16Feb2007

QC 2 16 07 FORTE 1

16Feb2007




HEAD 1



HEAD 2

Uniformity - Quantification

QC 1 13 09 ARGUS



FLOOD

UFOV

Full Report of Uniformity Analysis

NAME: QC 1 13 09 ARGUS ID:

DK DATE: 13Jan2009

UFOV

Integral Uniformity = 4.49%
Counts Location
Minimum 2331 (6, 18)
Maximum 2550 (30, 26)

Row Differential Uniformity = 3.28%
Column Differential Uniformity = 2.41%

	Diff.	Location
Max Row	158	(6, 18)
Max Col	120	(29, 26)

CFOV

Integral Uniformity = 3.70%
Counts Location
Minimum 2368 (31, 44)
Maximum 2550 (30, 26)

Row Differential Uniformity = 2.81%
Column Differential Uniformity = 2.41%


	Diff.	Location
Max Row	138	(26, 27)
Max Col	120	(29, 26)

CFOV

Integral Uniformity should be $< 5\%$ for 5M count extrinsic flood for camera following NEMA method for calculation. Refer to vendor specifications.

Uniformity Quantification

PHYSICS FORTE2, SLW



DET 1 LO INTR

UFOV



CFOV

Full Report of Uniformity Analysis

NAME: PHYSICS FORTE2, SLW ID: 1132009SW DATE: 13Jan2009

UFOV

Integral Uniformity = 2.76%

	Counts	Location
Minimum	2273	(8, 38)
Maximum	2402	(32, 47)

Row Differential Uniformity = 2.15%

Column Differential Uniformity = 2.36%

	Diff.	Location
Max Row	100	(27, 49)
Max Col	110	(8, 34)

CFOV

Integral Uniformity = 2.55%

	Counts	Location
Minimum	2273	(48, 24)
Maximum	2392	(38, 25)

Row Differential Uniformity = 1.89%

Column Differential Uniformity = 2.23%


	Diff.	Location
Max Row	88	(38, 37)
Max Col	104	(13, 23)

Uniformity – Not so Good

QC 7 8 09 FORTE 3

Full Report of Uniformity Analysis
NAME: QC 7 8 09 FORTE 3 ID:

- DATE: 08Jul2009



FLOOD HD1


UFOV

UFOV

Integral Uniformity = 7.16%
Counts Location
Minimum 2167 (8, 25)
Maximum 2501 (19, 24)

Row Differential Uniformity = 4.39%
Column Differential Uniformity = 4.03%

	Diff.	Location
Max Row	207	(12, 24)
Max Col	192	(16, 18)



CFOV

CFOV

Integral Uniformity = 5.64%
Counts Location
Minimum 2234 (12, 28)
Maximum 2501 (19, 24)


Row Differential Uniformity = 4.39%
Column Differential Uniformity = 2.73%

	Diff.	Location
Max Row	207	(12, 24)
Max Col	133	(19, 24)

Flood Images – Off Peak

CLASS IMAGES 4-23

23Apr2007



HD1 HIGH BY 7

CLASS IMAGES 4-23


23Apr2007



HD2 LOW BY 7

CLASS IMAGES 4-23


23Apr2007



HD1 HI BY 14

CLASS IMAGES 4-23


23Apr2007



HD2 LO BY 14

Low to High Count Rate Intrinsic Floods


19 kcps



DET 2 TC LO

04/16/09 17:05:49


79 kcps



DET 2 TC HIGH2

04/16/09 17:22:48

109 kcps
(too high)




DET 2 TC HIGH3

04/16/09 17:26:47

Routine QC – Spatial Resolution and Linearity

- Image bar pattern at least weekly to check spatial resolution and linearity
- May be done extrinsically using ^{57}Co or $^{99\text{m}}\text{Tc}$ sheet source and bar pattern placed on top of collimator
- May be done intrinsically using $^{99\text{m}}\text{Tc}$ point source at a distance, with bar pattern placed on top of crystal.
- 2.5 million counts adequate for routine QC

Spatial Resolution – Four Quadrant Bar Pattern



Intrinsic Spatial Resolution



Point Source ^{99m}Tc , 400-800 μCi

Collimator off,
bar pattern on
top of crystal

Extrinsic Spatial Resolution




Bar pattern on top of collimator, sheet source on top of bars


Routine QC – Spatial Resolution and Linearity

- Resolution - check that smallest resolvable bar pattern remains the same, no abrupt changes
- Linearity – check that lines on bar pattern do not appear significantly wavy and that there is no abrupt change

Bar pattern – intrinsic vs. extrinsic




Intrinsic – better resolution
than extrinsic




Extrinsic

Intrinsic Bar Pattern Tc and Thallium




^{99m}Tc




^{201}Tl

Intrinsic Thallium bar pattern, One Peak at a time




Lower energy peak only, 69 keV




Upper energy peak only, 167 keV –
Better resolution at higher energy

Extrinsic Bars – ^{99m}Tc and ^{57}Co



^{99m}Tc 140 keV γ




^{57}Co 122 keV γ

Tc resolution better due to higher energy gamma

Bar pattern – slight nonlinearity in corner



Bar Pattern with wrong collimator Medium Energy Collimator





SPECT QC – Center of Rotation

- Test or calibrate COR (center of rotation) corrections at least monthly
- Follow manufacturer's recommendations and instructions for testing/calibration procedures

Center of Rotation

Offsets between physical center of rotation and center of image matrix must be corrected for.

Pixel number recording signal
128x128 matrix



Center of Rotation

Date Jul 15, 2009

Rel Angle 90 deg

Collimator VXGP

X Errors

Detector1

Detector2

Mean 0.18 mm 0.26 mm

Max 0.45mm 0.55 mm

Min -0.23 mm -0.039 mm

Range 0.68 mm 0.59 mm

Y Errors

Detector1

Detector2

Max 0.49 mm 0.31 mm

Min -0.21 mm -0.42 mm


Range 0.7 mm 0.73 mm

Mean Y Difference 0.138 mm

Close

Print

Delete File



Files
CORQA0190.20090715173501
and
CORQA0290.20090715173501
PASS. The current calibration
files are good

/export/home/atlas/data/Calibra

Calibration Files

```
.  
90.20071005154245  
90.20070712112047  
180.20070712105620  
90.20070619094139  
180.20070619084821  
90.20070501115008
```

QA Files

```
.  
QA90.20090715173501  
QA90.20090522164038  
QA90.20090417170502  
QA90.20090317162025  
QA90.20090220164815  
QA90.20081219170942
```


Summary of Routine QC

- Check photopeak daily and adjust as needed.
- Check uniformity daily and take action if uniformity unacceptable.
- Check bar pattern for resolution and linearity at least weekly.
- Check/calibrate COR at least monthly if camera used for SPECT.

Accrediting Bodies and QC Recommendations

- Accreditation is required by some insurance companies for full reimbursement
- Accreditation in nuclear medicine is offered by:
 - American College of Radiology (ACR)
 - and
 - Intersocietal Commission for the Accreditation of Nuclear Medicine Laboratories (ICANL)

Accrediting Bodies and QC Recommendations

- ACR requires specific tests be done on a gamma camera and the images submitted for review by physicists
- ICANL does not require submission of physics tests or phantom images
- Both ACR and ICANL have recommendations for QC

Routine QC for Technologists – ACR Guidelines

- Intrinsic or System Uniformity – daily
- Intrinsic or System Resolution – weekly
- COR or Multiple Detector Registration Calibration/Test for SPECT systems – monthly
- High count floods for uniformity correction – as recommended by medical physicist
- Overall system performance for SPECT systems
- quarterly SPECT phantom, ^{99m}Tc at least semiannually, other radionuclides on alternate quarters

ICANL Guidelines for Gamma Camera QC

- Energy peaking – daily
- Intrinsic or extrinsic uniformity – daily
- Resolution and linearity (bar pattern)- weekly
- High count floods (≥ 30 M counts) – monthly or per manufacturer's recommendation
- Center of rotation – monthly
- Collimator integrity – annually
- Uniformity calibration – monthly or per manufacturer's recommendations
- Preventative Maintenance – every 6 months



SPECT Phantom

- Jaszczak Phantom used for many years for SPECT quality control.
- This phantom is approved by ACR for SPECT ACR Accreditation images



SPECT Phantom Imaging

- Deluxe version has spheres of diameters: 31.8, 25.4, 19.1, 15.9, 12.7, 9.5 mm
- Rods of diameters: 12.7, 11.1, 9.5, 7.9, 6.4 and 4.8 mm




SPECT phantom imaging procedure

- Make sure largest sphere lined up with largest rod section (rotate if needed)
- Fill phantom with ~20-25 mCi ^{99m}Tc for high res collimator (too much activity causes excessively high count rate and possible artifacts. Too little activity takes a long time to image). Keep count rate < 30kcps
- Center phantom in field of view


SPECT phantom imaging procedure

- ACR protocol is for 24 M total counts. Check count rate and adjust time per stop to achieve this
- Use 128 X 128 matrix, 120 or 128 views over 360 degrees
- Use a radius of rotation as close to 20 cm as possible
- For a large field of view camera, set the zoom between 1.33 and 1.46


SPECT phantom reconstructed slices



SPECT phantom reconstructed slices – no attenuation correction




Summed Slices



ACR accreditation scoring criteria (from ACR website)


- Phantom images are scored for:
 - Resolution – smallest size of rods visible
 - Contrast – number of spheres visible
 - Uniformity – look for ring type artifacts or other artifacts
- For specifics, refer to ACR website, since criteria vary according to type of collimator and which radionuclide used

Ring Artifacts




Ring artifacts
visible


Ring Artifacts



Ring Artifacts




Severe Ring Artifacts



SPECT Ring Artifacts

- Caused by non-uniformities such as:
 - Visible non-uniformities in flood image due to camera being off peak, PMT gain imbalance, or need for new correction tables
 - Shift in photopeak as camera head rotates
 - Collimator defect or damage (not visible in intrinsic flood image)
- Even small non-uniformities can cause ring artifacts

SPECT phantom planar resolution images



Static images of SPECT phantom standing on end on top of collimator. Provides a measure of planar system resolution with scatter.

Annual Gamma Camera Tests – ACR Guidelines

- Intrinsic and system uniformity
- Energy resolution
- Intrinsic or system spatial resolution
- High count rate performance
- Sensitivity
- System interlocks
- If camera used for SPECT, SPECT phantom images to evaluate tomographic uniformity, contrast and spatial resolution

Annual Gamma Camera Tests – Suggested Procedures

Intrinsic Uniformity:

- Acquire flood images with at least 5 M total counts
- If Thallium is used routinely, acquire Thallium uniformity image as well as ^{99m}Tc
- Acquire a high count rate image to test performance at high count rates



Annual Gamma Camera Tests – Suggested Procedures

System Uniformity:

- Acquire images with at least 5 M total counts
- Use ^{99m}Tc or ^{57}Co Sheet source
- Check collimators for defects
- Can check medium or high energy collimators for defects with ^{99m}Tc or ^{57}Co sheet source.

High Energy Collimator with ^{99m}Tc Sheet Source

Bubble



Annual Tests – Spatial Resolution

- Intrinsic spatial resolution should be checked, especially if not done routinely
- If four quadrant bar pattern used, rotate bar pattern to check smallest resolvable bars in both x and y direction

Annual Gamma Camera Tests – Suggested Procedures



Energy Resolution:

Defined as FWHM of photopeak divided by the mean energy.

Intrinsic energy resolution specified by manufacturer for ^{99m}Tc , typically 9-10%

Can estimate visually by observing photopeak and adjusting energy window

Energy Resolution



- Estimate ~ 9% energy resolution – photopeak width is approximate width of 9% window at half the peak height

Annual Gamma Camera Tests – Suggested Procedures

- Sensitivity:
 - Place ~1-2 mCi ^{99m}Tc and a small amount of water in a small plastic flat-bottomed vial. Record exact activity and time
 - Place on top of Styrofoam cup or similar object 10 cm from collimator face



Annual Gamma Camera Tests – Suggested Procedures


- Sensitivity:
 - Acquire 1 minute image, also a 1 min background image
 - Resulting counts/min per μCi can be compared with vendor specification for the collimator used. Also if > 1 camera head, sensitivity should be similar for all heads.



Unusual Artifacts

Tube-like Artifacts visible in Parathyroid study

NOT FOR DIAGNOSTIC PURPOSES



DET 1

04/30/09 10:20:22

Cause: ^{99m}Tc Aerosol from previous patient pulled into camera head by fans






Image made with no other source other than aerosol pulled into camera heads. Tc present behind the crystal. Patient breathed aerosol close to camera heads and leakage occurred around breathing apparatus.

Daily QC – Water filled sheet source with ^{99m}Tc MAA added

Morning QC looked bad



QC repeated with same source later in afternoon



Note “clumping” effect caused by using MAA. ^{99m}Tc MAA (normally used for lung perfusion imaging) was used by mistake. Should use ^{99m}Tc pertechnetate. Sheet source had to be emptied and rinsed

The End